

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011625\
 Data File : VU062833.D
 Acq On : 16 Jan 2025 21:00
 Operator : MD/SY
 Sample : VU0116WBL03
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK144

Quant Time: Jan 16 23:12:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 16 23:11:57 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	95140	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	88721	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	40862	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	23859	4.528	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.600%
7) Chloroethane-d5	1.904	69	20709	4.886	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.800%
11) 1,1-Dichloroethene-d2	2.557	65	12900	4.726	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.600%
20) 2-Butanone-d5	4.605	46	52978	49.194	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.380%
24) Chloroform-d	5.049	84	62986	4.858	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
26) 1,2-Dichloroethane-d4	5.689	65	35757	5.085	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.800%
32) Benzene-d6	5.715	84	107492	4.956	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
36) 1,2-Dichloropropane-d6	6.676	67	29415	4.892	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.800%
41) Toluene-d8	7.888	98	104409	4.941	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.800%
43) trans-1,3-Dichloroprop...	8.171	79	14839	4.951	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.000%
46) 2-Hexanone-d5	8.621	63	43802	46.465	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.920%
56) 1,1,2,2-Tetrachloroeth...	10.743	84	25424	5.049	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.000%
66) 1,2-Dichlorobenzene-d4	12.184	152	36006	5.296	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011625\
Data File : VU062833.D
Acq On : 16 Jan 2025 21:00
Operator : MD/SY
Sample : VU0116WBL03
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK144

Quant Time: Jan 16 23:12:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 16 23:11:57 2025
Response via : Initial Calibration

