

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011625\
 Data File : VU062852.D
 Acq On : 17 Jan 2025 04:48
 Operator : MD/SY
 Sample : Q1072-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCQ89

Quant Time: Jan 17 05:14:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 16 23:11:57 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	86852	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.406	117	83036	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	38391	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18760	3.900	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.000%
7) Chloroethane-d5	1.901	69	17067	4.411	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.200%
11) 1,1-Dichloroethene-d2	2.557	65	10908	4.377	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.600%
20) 2-Butanone-d5	4.605	46	47083	47.892	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.780%
24) Chloroform-d	5.049	84	52980	4.477	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
26) 1,2-Dichloroethane-d4	5.689	65	31206	4.861	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.200%
32) Benzene-d6	5.714	84	88544	4.362	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.200%
36) 1,2-Dichloropropane-d6	6.676	67	24652	4.380	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.600%
41) Toluene-d8	7.888	98	84292	4.263	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	8.171	79	12238	4.362	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.200%
46) 2-Hexanone-d5	8.618	63	36693	41.588	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.180%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	21002	4.456	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.200%
66) 1,2-Dichlorobenzene-d4	12.184	152	30281	4.741	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011625\
Data File : VU062852.D
Acq On : 17 Jan 2025 04:48
Operator : MD/SY
Sample : Q1072-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCQ89

Quant Time: Jan 17 05:14:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 16 23:11:57 2025
Response via : Initial Calibration

