

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011720\
 Data File : VU036509.D
 Acq On : 17 Jan 2020 11:10
 Operator : JC/MD
 Sample : VU0117MBL01
 Misc : 5.00G/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK44

Quant Time: Jan 17 14:10:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM011420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 17 14:05:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	511427	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	482141	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	201595	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	154231	46.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.02%
7) Chloroethane-d5	1.93	69	139403	48.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.14%
11) 1,1-Dichloroethene-d2	2.59	63	228085	35.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.64%
21) 2-Butanone-d5	4.68	46	303379	106.42	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.42%
24) Chloroform-d	5.11	84	332936	50.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.88%
26) 1,2-Dichloroethane-d4	5.75	65	227845	51.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.84%
32) Benzene-d6	5.77	84	642476	50.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.90%
36) 1,2-Dichloropropane-d6	6.73	67	209405	51.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.12%
41) Toluene-d8	7.93	98	591704	49.61	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.22%
43) trans-1,3-Dichloropropene-	8.21	79	107007	48.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.88%
47) 2-Hexanone-d5	8.67	63	221575	105.44	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.44%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	314625	51.01	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.02%
64) 1,2-Dichlorobenzene-d4	12.21	152	210023	52.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.78%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011720\
Data File : VU036509.D
Acq On : 17 Jan 2020 11:10
Operator : JC/MD
Sample : VU0117MBL01
Misc : 5.00G/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK44

Quant Time: Jan 17 14:10:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM011420WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 17 14:05:23 2020
Response via : Initial Calibration

