

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011723\
 Data File : VU052714.D
 Acq On : 17 Jan 2023 12:12
 Operator : JC/MD
 Sample : VU0117WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK171

Quant Time: Jan 18 01:16:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 18 01:16:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	130372	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	126438	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	56489	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	50873	4.442	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.800%
7) Chloroethane-d5	1.916	69	45103	4.834	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.600%
11) 1,1-Dichloroethene-d2	2.565	65	25683	5.092	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.800%
20) 2-Butanone-d5	4.642	46	148787	48.584	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.160%
24) Chloroform-d	5.057	84	109888	5.143	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
26) 1,2-Dichloroethane-d4	5.700	65	58467	5.329	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.600%
32) Benzene-d6	5.723	84	219550	5.357	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.200%
36) 1,2-Dichloropropane-d6	6.687	67	71032	5.201	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.000%
41) Toluene-d8	7.893	98	180282	4.990	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.800%
43) trans-1,3-Dichloroprop...	8.176	79	25092	4.941	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.800%
46) 2-Hexanone-d5	8.632	63	102387	48.533	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.060%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	52437	5.015	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.400%
66) 1,2-Dichlorobenzene-d4	12.189	152	60599	5.893	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011723\
Data File : VU052714.D
Acq On : 17 Jan 2023 12:12
Operator : JC/MD
Sample : VU0117WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK171

Quant Time: Jan 18 01:16:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 18 01:16:13 2023
Response via : Initial Calibration

