

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011725\
 Data File : VU062869.D
 Acq On : 17 Jan 2025 14:13
 Operator : MD/SY
 Sample : Q1057-21
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jan 17 21:59:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 17 21:53:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	86488	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	82774	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	39237	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	19779	4.129	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.600%
7) Chloroethane-d5	1.901	69	18279	4.744	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.800%
11) 1,1-Dichloroethene-d2	2.557	65	11031	4.445	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.000%
20) 2-Butanone-d5	4.605	46	50302	51.381	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.760%
24) Chloroform-d	5.049	84	57844	4.908	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	5.689	65	34143	5.341	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
32) Benzene-d6	5.714	84	95475	4.719	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
36) 1,2-Dichloropropane-d6	6.679	67	26971	4.808	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.200%
41) Toluene-d8	7.888	98	86789	4.403	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
43) trans-1,3-Dichloroprop...	8.171	79	13291	4.753	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.000%
46) 2-Hexanone-d5	8.621	63	41924	47.668	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.340%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	24039	5.117	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.400%
66) 1,2-Dichlorobenzene-d4	12.184	152	33870	5.189	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011725\
Data File : VU062869.D
Acq On : 17 Jan 2025 14:13
Operator : MD/SY
Sample : Q1057-21
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jan 17 21:59:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 17 21:53:56 2025
Response via : Initial Calibration

