

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011725\
 Data File : VU062903.D
 Acq On : 18 Jan 2025 04:31
 Operator : MD/SY
 Sample : Q1126-04
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COAH7

Quant Time: Jan 18 04:59:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 17 23:41:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	86819	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	79797	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	36985	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18297	3.805	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.200%
7) Chloroethane-d5	1.904	69	17046	4.407	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.200%
11) 1,1-Dichloroethene-d2	2.557	65	9125	3.663	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.200%
20) 2-Butanone-d5	4.605	46	54677	55.637	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.280%
24) Chloroform-d	5.049	84	54695	4.623	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
26) 1,2-Dichloroethane-d4	5.689	65	32343	5.040	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.714	84	87726	4.497	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.000%
36) 1,2-Dichloropropane-d6	6.679	67	25532	4.721	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.888	98	78960	4.155	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.000%
43) trans-1,3-Dichloroprop...	8.171	79	11773	4.367	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.400%
46) 2-Hexanone-d5	8.621	63	41791	49.289	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.580%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	22944	5.066	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.400%
66) 1,2-Dichlorobenzene-d4	12.183	152	30780	5.002	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%
Target Compounds					Qvalue	
12) 1,1-Dichloroethene	2.570	96	1045	0.194	ug/L #	1
29) 1,1,1-Trichloroethane	5.303	97	10644	0.879	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011725\
Data File : VU062903.D
Acq On : 18 Jan 2025 04:31
Operator : MD/SY
Sample : Q1126-04
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AH7

Quant Time: Jan 18 04:59:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 17 23:41:19 2025
Response via : Initial Calibration

