

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU011819\
 Data File : VU029215.D
 Acq On : 18 Jan 2019 10:32
 Operator : JC/SP
 Sample : VU0118WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK50

Quant Time: Jan 18 23:13:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 18 23:01:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	119829	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	117297	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	50397	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	36114	5.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.60%
7) Chloroethane-d5	1.68	69	30993	5.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.80%
11) 1,1-Dichloroethene-d2	2.27	63	51111	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	4.22	46	105786	63.12	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	126.24%
24) Chloroform-d	4.65	84	71584	5.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.80%
26) 1,2-Dichloroethane-d4	5.31	65	36676	5.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.40%
32) Benzene-d6	5.34	84	148422	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.20%
36) 1,2-Dichloropropane-d6	6.33	67	49998	5.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.00%
41) Toluene-d8	7.56	98	126696	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
43) trans-1,3-Dichloropropene-	7.85	79	18474	5.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	112.20%
46) 2-Hexanone-d5	8.31	63	90388	60.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	120.30%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	39058	5.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	45813	5.43	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU011819\
Data File : VU029215.D
Acq On : 18 Jan 2019 10:32
Operator : JC/SP
Sample : VU0118WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK50

Quant Time: Jan 18 23:13:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 18 23:01:06 2019
Response via : Initial Calibration

