

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU011819\
 Data File : VU029216.D
 Acq On : 18 Jan 2019 11:03
 Operator : JC/SP
 Sample : K1155-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEJG2

Quant Time: Jan 18 23:15:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 18 23:01:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	115163	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	111488	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	48876	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	34316	5.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.40%
7) Chloroethane-d5	1.68	69	29386	5.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.20%
11) 1,1-Dichloroethene-d2	2.27	63	49584	3.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.00%
20) 2-Butanone-d5	4.22	46	103686	64.37	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	128.74%
24) Chloroform-d	4.65	84	69300	5.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.60%
26) 1,2-Dichloroethane-d4	5.31	65	34799	5.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.80%
32) Benzene-d6	5.34	84	142647	5.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.40%
36) 1,2-Dichloropropane-d6	6.33	67	50094	6.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	120.20%
41) Toluene-d8	7.56	98	122226	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
43) trans-1,3-Dichloropropene-	7.85	79	18244	5.82	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	116.40%
46) 2-Hexanone-d5	8.32	63	83685	58.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	117.18%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	37358	5.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	44015	5.38	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU011819\
Data File : VU029216.D
Acq On : 18 Jan 2019 11:03
Operator : JC/SP
Sample : K1155-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEJG2

Quant Time: Jan 18 23:15:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR122818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 18 23:01:06 2019
Response via : Initial Calibration

