

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_U\DATA\VU011819\  
 Data File : VU029218.D  
 Acq On : 18 Jan 2019 11:59  
 Operator : JC/SP  
 Sample : K1155-03  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VHBLK01

Quant Time: Jan 18 23:21:33 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR122818WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Fri Jan 18 23:01:06 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.88  | 114  | 110518   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.09  | 117  | 111807   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.48 | 152  | 47911    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.40   | 65    | 34227    | 5.79     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 115.80% |
| 7) Chloroethane-d5             | 1.68   | 69    | 27620    | 5.35     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 107.00% |
| 11) 1,1-Dichloroethene-d2      | 2.27   | 63    | 48186    | 3.85     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 77.00%  |
| 20) 2-Butanone-d5              | 4.21   | 46    | 99563    | 64.41    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 128.82% |
| 24) Chloroform-d               | 4.65   | 84    | 66805    | 5.81     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 116.20% |
| 26) 1,2-Dichloroethane-d4      | 5.31   | 65    | 34314    | 5.85     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 117.00% |
| 32) Benzene-d6                 | 5.34   | 84    | 135542   | 5.19     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 103.80% |
| 36) 1,2-Dichloropropane-d6     | 6.33   | 67    | 47793    | 5.72     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 114.40% |
| 41) Toluene-d8                 | 7.56   | 98    | 115990   | 4.79     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 95.80%  |
| 43) trans-1,3-Dichloropropene- | 7.85   | 79    | 16156    | 5.14     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 102.80% |
| 46) 2-Hexanone-d5              | 8.31   | 63    | 81091    | 56.61    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 113.22% |
| 57) 1,1,2,2-Tetrachloroethane- | 10.43  | 84    | 36273    | 5.45     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 109.00% |
| 64) 1,2-Dichlorobenzene-d4     | 11.86  | 152   | 41928    | 5.23     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 104.60% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_U\DATA\VU011819\  
Data File : VU029218.D  
Acq On : 18 Jan 2019 11:59  
Operator : JC/SP  
Sample : K1155-03  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VHBLK01

Quant Time: Jan 18 23:21:33 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR122818WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Fri Jan 18 23:01:06 2019  
Response via : Initial Calibration

