

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012021\
 Data File : VU042118.D
 Acq On : 20 Jan 2021 20:12
 Operator : SY/MD
 Sample : M1180-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F1E05

Quant Time: Jan 21 04:27:16 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 21 04:09:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	118393	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	110112	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	58146	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	34402	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	1.919	69	28518	4.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.80%
11) 1,1-Dichloroethene-d2	2.568	63	52931	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	4.662	46	115226	41.35	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.70%
24) Chloroform-d	5.073	84	71227	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.40%
26) 1,2-Dichloroethane-d4	5.713	65	43728	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
32) Benzene-d6	5.735	84	136503	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.700	67	44164	4.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.20%
41) Toluene-d8	7.906	98	124821	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
43) trans-1,3-Dichloroprop...	8.189	79	21302	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
46) 2-Hexanone-d5	8.645	63	96306	47.05	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.10%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	37379	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.00%
64) 1,2-Dichlorobenzene-d4	12.201	152	52530	5.03	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.60%
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	4.674	96	14117	1.55	ug/L	96
33) Benzene	5.777	78	1083	0.03	ug/L	100
40) 4-Methyl-2-pentanone	7.806	43	3357	0.46	ug/L #	89
42) Toluene	7.976	91	2092	0.06	ug/L	77
55) Styrene	10.121	104	1127	0.04	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012021\
Data File : VU042118.D
Acq On : 20 Jan 2021 20:12
Operator : SY/MD
Sample : M1180-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F1E05

Quant Time: Jan 21 04:27:16 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 21 04:09:50 2021
Response via : Initial Calibration

