

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012023\
 Data File : VU052752.D
 Acq On : 20 Jan 2023 10:59
 Operator : JC/MD
 Sample : VU0120WBL06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0120WBL06

Quant Time: Jan 21 01:00:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U011923W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 20 01:04:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	200693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	332682	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	302076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.809	152	136277	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	160853	55.652	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.300%
35) Dibromofluoromethane	5.285	113	123214	53.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.380%
50) Toluene-d8	7.893	98	437085	50.313	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.620%
62) 4-Bromofluorobenzene	10.629	95	146174	46.062	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012023\
Data File : VU052752.D
Acq On : 20 Jan 2023 10:59
Operator : JC/MD
Sample : VU0120WBL06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0120WBL06

Quant Time: Jan 21 01:00:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U011923W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 20 01:04:16 2023
Response via : Initial Calibration

