

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012025\
 Data File : VU062908.D
 Acq On : 20 Jan 2025 10:40
 Operator : MD/SY
 Sample : VU0120WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK147

Quant Time: Jan 21 02:39:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 21 02:39:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	89812	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	83006	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	38339	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	19664	3.953	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.000%
7) Chloroethane-d5	1.904	69	18318	4.578	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.600%
11) 1,1-Dichloroethene-d2	2.557	65	10570	4.102	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.000%
20) 2-Butanone-d5	4.605	46	56368	55.446	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.900%
24) Chloroform-d	5.049	84	58849	4.809	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%
26) 1,2-Dichloroethane-d4	5.689	65	34086	5.135	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.600%
32) Benzene-d6	5.715	84	91425	4.506	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
36) 1,2-Dichloropropane-d6	6.679	67	27795	4.941	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.800%
41) Toluene-d8	7.888	98	82934	4.195	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	8.171	79	13049	4.653	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.000%
46) 2-Hexanone-d5	8.618	63	42624	48.328	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.660%
56) 1,1,2,2-Tetrachloroeth...	10.743	84	23997	5.093	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.800%
66) 1,2-Dichlorobenzene-d4	12.184	152	31587	4.952	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012025\
Data File : VU062908.D
Acq On : 20 Jan 2025 10:40
Operator : MD/SY
Sample : VU0120WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK147

Quant Time: Jan 21 02:39:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 21 02:39:03 2025
Response via : Initial Calibration

