

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012025\
 Data File : VU062920.D
 Acq On : 20 Jan 2025 15:30
 Operator : MD/SY
 Sample : Q1086-12DL 20X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COAFODL

Quant Time: Jan 21 02:42:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 21 02:39:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	92816	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.405	117	88970	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	41748	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	17793	3.461	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.200%
7) Chloroethane-d5	1.904	69	17057	4.125	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.400%
11) 1,1-Dichloroethene-d2	2.554	65	9824	3.689	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.800%
20) 2-Butanone-d5	4.605	46	59402	56.540	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.080%
24) Chloroform-d	5.049	84	57961	4.583	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.689	65	34239	4.991	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.800%
32) Benzene-d6	5.714	84	90811	4.176	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
36) 1,2-Dichloropropane-d6	6.679	67	28235	4.682	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.600%
41) Toluene-d8	7.888	98	80499	3.799	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.000%
43) trans-1,3-Dichloroprop...	8.171	79	13073	4.349	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.621	63	45459	48.088	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.180%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	24791	4.909	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.200%
66) 1,2-Dichlorobenzene-d4	12.184	152	33610	4.839	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.800%
Target Compounds						
19) 1,1-Dichloroethane	3.853	63	9868	0.894	ug/L	94
29) 1,1,1-Trichloroethane	5.303	97	128464	9.519	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012025\
Data File : VU062920.D
Acq On : 20 Jan 2025 15:30
Operator : MD/SY
Sample : Q1086-12DL 20X
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AF0DL

Quant Time: Jan 21 02:42:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 21 02:39:03 2025
Response via : Initial Calibration

