

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012119\
 Data File : VU029246.D
 Acq On : 21 Jan 2019 14:08
 Operator : JC/SP
 Sample : PB116506ZHE#06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB116506ZHE#06

Quant Time: Jan 22 01:56:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	112990	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	197582	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	192025	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	72594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	77834	55.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.44%	
35) Dibromofluoromethane	4.88	113	67319	52.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.12%	
50) Toluene-d8	7.56	98	260323	51.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.96%	
62) 4-Bromofluorobenzene	10.30	95	81400	44.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.58%	
Target Compounds						
16) Acetone	2.32	43	4619	5.249	ug/l	93
20) Methylene Chloride	2.70	84	7804	5.224	ug/l	98
43) Isopropyl Acetate	5.55	43	8230	2.475	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU012119\
Data File : VU029246.D
Acq On : 21 Jan 2019 14:08
Operator : JC/SP
Sample : PB116506ZHE#06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB116506ZHE#06

Quant Time: Jan 22 01:56:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

