

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012119\
 Data File : VU029251.D
 Acq On : 21 Jan 2019 16:08
 Operator : JC/SP
 Sample : K1013-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PL-01-011819-B

Quant Time: Jan 22 02:47:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	112407	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	193749	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	189558	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	71471	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	76817	55.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.56%	
35) Dibromofluoromethane	4.88	113	66914	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
50) Toluene-d8	7.56	98	257875	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
62) 4-Bromofluorobenzene	10.30	95	81296	45.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.24%	
Target Compounds						
16) Acetone	2.32	43	7418	8.474	ug/l	97
20) Methylene Chloride	2.70	84	140076	94.248	ug/l	98
43) Isopropyl Acetate	5.55	43	6823	2.092	ug/l	95
95) Naphthalene	13.75	128	18062	3.266	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU012119\
Data File : VU029251.D
Acq On : 21 Jan 2019 16:08
Operator : JC/SP
Sample : K1013-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PL-01-011819-B

Quant Time: Jan 22 02:47:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

