

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012119\
 Data File : VU029252.D
 Acq On : 21 Jan 2019 16:32
 Operator : JC/SP
 Sample : K1013-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PL-01-011819-C

Quant Time: Jan 22 02:50:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	110441	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	193964	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	185493	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	76420	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	75575	55.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.72%	
35) Dibromofluoromethane	4.88	113	65879	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
50) Toluene-d8	7.56	98	254696	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	10.30	95	80243	44.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.96%	
Target Compounds						
16) Acetone	2.32	43	6056	7.041	ug/l	98
20) Methylene Chloride	2.70	84	75139	51.456	ug/l	96
32) 1,1,1-Trichloroethane	4.91	97	3050	1.616	ug/l	93
43) Isopropyl Acetate	5.55	43	7026	2.152	ug/l #	93
84) 1,2,4-Trimethylbenzene	11.15	105	6251	1.325	ug/l	96
95) Naphthalene	13.75	128	9240	1.562	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU012119\
Data File : VU029252.D
Acq On : 21 Jan 2019 16:32
Operator : JC/SP
Sample : K1013-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PL-01-011819-C

Quant Time: Jan 22 02:50:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

