

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012119\
 Data File : VU029254.D
 Acq On : 21 Jan 2019 17:20
 Operator : JC/SP
 Sample : K1015-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SU-01-011719-B

Quant Time: Jan 22 02:55:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	110177	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	195997	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	184646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	75978	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.31	65	77012	56.54	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	113.08%	
35) Dibromofluoromethane	4.88	113	67310	52.98	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	105.96%	
50) Toluene-d8	7.56	98	255041	50.85	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	10.30	95	80407	44.61	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.22%	

Target Compounds

						Qvalue
16) Acetone	2.32	43	25030	29.172	ug/l	99
20) Methylene Chloride	2.70	84	106796	73.311	ug/l	96
43) Isopropyl Acetate	5.55	43	6811	2.065	ug/l #	87
95) Naphthalene	13.74	128	104983	17.856	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012119\
 Data File : VU029254.D
 Acq On : 21 Jan 2019 17:20
 Operator : JC/SP
 Sample : K1015-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampled :
 SU-01-011719-B

Quant Time: Jan 22 02:55:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

