

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012121\
 Data File : VU042145.D
 Acq On : 21 Jan 2021 12:59
 Operator : SY/MD
 Sample : M1170-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFT95

Manual Integrations
 APPROVED

SAM
 1/25/2021 8:03:27 AM

Quant Time: Jan 22 02:46:52 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 22 02:43:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	113545	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	105585	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	54018	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	31770	4.34	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.80%
7) Chloroethane-d5	1.919	69	26333	4.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.20%
11) 1,1-Dichloroethene-d2	2.571	63	55429	3.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.60%
20) 2-Butanone-d5	4.649	46	138861	51.96	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.92%
24) Chloroform-d	5.073	84	70073	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.60%
26) 1,2-Dichloroethane-d4	5.710	65	44057	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.735	84	132838	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.40%
36) 1,2-Dichloropropane-d6	6.700	67	43333	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.906	98	119994	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
43) trans-1,3-Dichloroprop...	8.185	79	20207	4.48	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.60%
46) 2-Hexanone-d5	8.645	63	98647	50.26	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.52%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	36017	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	12.201	152	47689	4.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.20%
Target Compounds					Qvalue	
13) Acetone	2.665	43	4792m	2.69	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012121\
Data File : VU042145.D
Acq On : 21 Jan 2021 12:59
Operator : SY/MD
Sample : M1170-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1


Instrument :
MSVOA_U
ClientSampleId :
BFT95

Manual Integrations
APPROVED

SAM
1/25/2021 8:03:27 AM


Quant Time: Jan 22 02:46:52 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 22 02:43:44 2021
Response via : Initial Calibration





#13
 Acetone
 Concen: 2.69 ug/L m
 RT: 2.665 min Scan# 412
 Delta R.T. 0.006 min
 Lab File: VU042145.D
 Acq: 21 Jan 2021 12:59

Instrument :
 MSVOA_U
 ClientSampleId :
 BFT95



Tgt Ion: 43	Resp: 4792
Ion Ratio	Lower Upper
43	100
58	15.1 0.0 55.8

Manual Integrations
 APPROVED
 SAM
 1/25/2021 8:03:27 AM

