


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012122\
Data File : VU046820.D
Acq On : 21 Jan 2022 11:29
Operator : SY/MD
Sample : VU0121WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK053

Manual IntegrationsAPPROVED

Quant Time: Jan 24 00:24:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jan 24 00:23:26 2022
Response via : Initial Calibration

Reviewed By :John Carlone 01/24/2022
Supervised By :Mahesh Dadoda 01/24/2022






Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012122\
Data File : VU046820.D
Acq On : 21 Jan 2022 11:29
Operator : SY/MD
Sample : VU0121WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK053

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 01/24/2022
Supervised By :Mahesh Dadoda 01/24/2022

Quant Time: Jan 24 00:24:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jan 24 00:23:26 2022
Response via : Initial Calibration



TIC: VU046820.D\data.ms

(13) Acetone (T)

2.681min (-0.000) 2.31 ug/L

response 3105

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	35.80	36.23
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012122\
 Data File : VU046820.D
 Acq On : 21 Jan 2022 11:29
 Operator : SY/MD
 Sample : VU0121WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK053

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 01/24/2022
 Supervised By :Mahesh Dadoda 01/24/2022

Quant Time: Jan 24 00:24:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 24 00:23:26 2022
 Response via : Initial Calibration



TIC: VU046820.D\data.ms

(13) Acetone (T)

2.681min (-0.000) 4.00 ug/L m

response 5369

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	35.80	20.95
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012122\
 Data File : VU046820.D
 Acq On : 21 Jan 2022 11:29
 Operator : SY/MD
 Sample : VU0121WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK053

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 01/24/2022
 Supervised By :Mahesh Dadoda 01/24/2022

Quant Time: Jan 24 00:24:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 24 00:23:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	98362	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	105789	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	52781	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	31683	3.801	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	76.000%		
7) Chloroethane-d5	1.919	69	26988	4.110	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	82.200%		
11) 1,1-Dichloroethene-d2	2.575	65	12998	3.920	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	78.400%		
20) 2-Butanone-d5	4.655	46	133016	53.870	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	107.740%		
24) Chloroform-d	5.067	84	63727	4.355	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	87.200%		
26) 1,2-Dichloroethane-d4	5.706	65	38299	4.807	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	96.200%		
32) Benzene-d6	5.732	84	128492	4.133	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	82.600%		
36) 1,2-Dichloropropane-d6	6.693	67	41546	4.450	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	89.000%		
41) Toluene-d8	7.902	98	106676	3.748	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	75.000%		
43) trans-1,3-Dichloroprop...	8.182	79	16904	4.123	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	82.400%		
46) 2-Hexanone-d5	8.639	63	124604	51.330	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	102.660%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	45738	4.880	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	97.600%		
66) 1,2-Dichlorobenzene-d4	12.195	152	42303	4.274	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	85.400%		
Target Compounds					Qvalue	
13) Acetone	2.681	43	5369m	3.998	ug/L	
42) Toluene	7.973	91	10804	0.363	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed