

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012122\
 Data File : VU046821.D
 Acq On : 21 Jan 2022 12:14
 Operator : SY/MD
 Sample : N1223-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YAXB2

Quant Time: Jan 24 00:24:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 24 00:23:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	92816	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	101833	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	50666	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	31030	3.945	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.000%
7) Chloroethane-d5	1.919	69	26889	4.339	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.800%
11) 1,1-Dichloroethene-d2	2.575	65	13284	4.246	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.000%
20) 2-Butanone-d5	4.655	46	135929	58.340	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.680%
24) Chloroform-d	5.067	84	61233	4.435	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
26) 1,2-Dichloroethane-d4	5.707	65	37648	5.007	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
32) Benzene-d6	5.732	84	123968	4.142	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.800%
36) 1,2-Dichloropropane-d6	6.694	67	40226	4.476	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.600%
41) Toluene-d8	7.903	98	100799	3.680	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.600%
43) trans-1,3-Dichloroprop...	8.182	79	15829	4.011	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.642	63	116546	49.876	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.760%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	44102	4.888	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	41855	4.405	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012122\
Data File : VU046821.D
Acq On : 21 Jan 2022 12:14
Operator : SY/MD
Sample : N1223-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YAXB2

Quant Time: Jan 24 00:24:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jan 24 00:23:26 2022
Response via : Initial Calibration

