

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012123\
 Data File : VU052763.D
 Acq On : 20 Jan 2023 16:18
 Operator : JC/MD
 Sample : VU0121WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK172

Quant Time: Jan 21 01:56:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 21 01:44:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	135171	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	132814	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	62929	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	45555	3.836	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.800%
7) Chloroethane-d5	1.916	69	39912	4.125	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.600%
11) 1,1-Dichloroethene-d2	2.565	65	22369	4.277	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.600%
20) 2-Butanone-d5	4.652	46	171774	54.099	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.200%
24) Chloroform-d	5.057	84	101069	4.563	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
26) 1,2-Dichloroethane-d4	5.697	65	58583	5.150	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	5.723	84	203126	4.719	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
36) 1,2-Dichloropropane-d6	6.687	67	64055	4.465	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.400%
41) Toluene-d8	7.893	98	172429	4.543	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	8.176	79	26120	4.896	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.000%
46) 2-Hexanone-d5	8.632	63	129169	58.289	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.580%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	56484	5.143	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.800%
66) 1,2-Dichlorobenzene-d4	12.189	152	62722	5.475	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.600%
Target Compounds						Qvalue
16) Methylene chloride	3.044	84	6735	0.432	ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012123\
Data File : VU052763.D
Acq On : 20 Jan 2023 16:18
Operator : JC/MD
Sample : VU0121WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK172

Quant Time: Jan 21 01:56:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 21 01:44:26 2023
Response via : Initial Calibration

