

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012125\
 Data File : VU062944.D
 Acq On : 21 Jan 2025 16:11
 Operator : MD/SY
 Sample : VIBLK199
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK199

Quant Time: Jan 22 02:24:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 22 02:20:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	89424	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.406	117	71958	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	39489	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	13729	2.772	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	55.400%
7) Chloroethane-d5	1.904	69	12494	3.136	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	62.800%#
11) 1,1-Dichloroethene-d2	2.557	65	9288	3.620	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.400%
20) 2-Butanone-d5	4.605	46	64021	63.248	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	126.500%
24) Chloroform-d	5.049	84	49435	4.057	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%
26) 1,2-Dichloroethane-d4	5.689	65	27698	4.191	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.800%
32) Benzene-d6	5.714	84	69955	3.977	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.600%
36) 1,2-Dichloropropane-d6	6.679	67	27978	5.737	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.800%
41) Toluene-d8	7.888	98	64992	3.793	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.800%
43) trans-1,3-Dichloroprop...	8.174	79	10220	4.204	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.000%
46) 2-Hexanone-d5	8.618	63	32453	42.446	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.900%
56) 1,1,2,2-Tetrachloroeth...	10.743	84	20174	4.939	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.800%
66) 1,2-Dichlorobenzene-d4	12.184	152	30070	4.577	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012125\
Data File : VU062944.D
Acq On : 21 Jan 2025 16:11
Operator : MD/SY
Sample : VIBLK199
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK199

Quant Time: Jan 22 02:24:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 22 02:20:07 2025
Response via : Initial Calibration

