

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012221\
 Data File : VU042165.D
 Acq On : 22 Jan 2021 13:21
 Operator : SY/MD
 Sample : M1222-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F1E72

Quant Time: Jan 23 06:20:10 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012021WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jan 23 06:18:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	101843	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	95438	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	48529	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.601	65	26915	4.10	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.00%
7) Chloroethane-d5	1.916	69	23937	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.40%
11) 1,1-Dichloroethene-d2	2.572	63	47699	3.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.00%
20) 2-Butanone-d5	4.655	46	127405	53.15	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.30%
24) Chloroform-d	5.073	84	64293	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
26) 1,2-Dichloroethane-d4	5.713	65	40457	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.735	84	120094	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.40%
36) 1,2-Dichloropropane-d6	6.700	67	40136	4.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.80%
41) Toluene-d8	7.906	98	107660	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
43) trans-1,3-Dichloroprop...	8.185	79	18611	4.57	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.40%
46) 2-Hexanone-d5	8.645	63	89723	50.57	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.14%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	33654	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.60%
64) 1,2-Dichlorobenzene-d4	12.201	152	44361	5.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012221\
Data File : VU042165.D
Acq On : 22 Jan 2021 13:21
Operator : SY/MD
Sample : M1222-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F1E72

Quant Time: Jan 23 06:20:10 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012021WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jan 23 06:18:59 2021
Response via : Initial Calibration

