

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012225\
 Data File : VU062959.D
 Acq On : 22 Jan 2025 12:59
 Operator : MD/SY
 Sample : Q1126-05
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jan 23 00:45:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 23 00:43:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	84188	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	79838	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	37589	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	14458	3.101	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.000%
7) Chloroethane-d5	1.904	69	13894	3.704	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.000%
11) 1,1-Dichloroethene-d2	2.560	65	9661	4.000	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.000%
20) 2-Butanone-d5	4.615	46	44201	46.383	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.760%
24) Chloroform-d	5.049	84	51117	4.456	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
26) 1,2-Dichloroethane-d4	5.689	65	29608	4.758	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	5.714	84	83852	4.297	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
36) 1,2-Dichloropropane-d6	6.679	67	23709	4.382	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.600%
41) Toluene-d8	7.891	98	75475	3.970	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.400%
43) trans-1,3-Dichloroprop...	8.174	79	11384	4.221	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.400%
46) 2-Hexanone-d5	8.624	63	32888	38.769	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.540%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	19777	4.364	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.200%
66) 1,2-Dichlorobenzene-d4	12.183	152	29079	4.650	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012225\
Data File : VU062959.D
Acq On : 22 Jan 2025 12:59
Operator : MD/SY
Sample : Q1126-05
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jan 23 00:45:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 23 00:43:08 2025
Response via : Initial Calibration

