

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012320\
 Data File : VU036536.D
 Acq On : 23 Jan 2020 13:04
 Operator : JC/MD
 Sample : L1134-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WS

Quant Time: Jan 27 09:23:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U012220DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jan 23 08:59:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	6.15	96	36217	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.66	95	12649	0.93	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.00%	
68) 1,2-Dichlorobenzene-d4	12.22	152	13102	0.94	ug/l	0.00
Spiked Amount 1.000			Recovery	=	94.00%	
Target Compounds						Ovalue
27) Chloroform	5.13	83	236302	13.954	ug/l	100
44) Bromodichloromethane	7.14	83	529755	42.526	ug/l	100
55) Dibromochloromethane	8.84	129	417619	48.545	ug/l	100
67) Bromoform	10.32	173	127226	25.830	ug/l	100
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU012320\
Data File : VU036536.D
Acq On : 23 Jan 2020 13:04
Operator : JC/MD
Sample : L1134-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WS

Quant Time: Jan 27 09:23:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U012220DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jan 23 08:59:25 2020
Response via : Initial Calibration

