

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012320\
 Data File : VU036539.D
 Acq On : 23 Jan 2020 14:34
 Operator : JC/MD
 Sample : L1134-08DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-UNRVOA-WSDL

Quant Time: Jan 24 02:35:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U012220DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jan 23 08:59:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	6.15	96	37414	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.66	95	13499	0.96	ug/l	0.00
Spiked Amount 1.000			Recovery	=	96.00%	
68) 1,2-Dichlorobenzene-d4	12.21	152	14210	0.98	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.00%	
Target Compounds						
					Ovalue	
3) Chloromethane	1.54	50	54938	4.143	ug/l	97
6) Chloroethane	1.95	64	42793	4.810	ug/l	99
17) 1,1-Dichloroethane	3.91	63	26031	1.472	ug/l	98
21) 2,2-Dichloropropane	4.71	77	16541	1.096	ug/l	94
25) Methyl tert-Butyl Ether	3.42	73	129216	5.012	ug/l	97
26) Bromochloromethane	5.02	128	4700	0.964	ug/l	93
29) 1,1-Dichloropropene	5.57	75	36744	2.567	ug/l	100
43) Dibromomethane	6.96	93	4625	0.850	ug/l	97
50) t-1,3-Dichloropropene	8.24	75	40063	3.124	ug/l	96
51) cis-1,3-Dichloropropene	7.64	75	49395	3.178	ug/l	98
53) 1,3-Dichloropropane	8.61	76	12407	0.938	ug/l	100
60) 1,1,1,2-Tetrachloroethane	9.56	131	8858	0.941	ug/l	99
69) Isopropylbenzene	10.51	105	141065	2.836	ug/l	100
70) 1,1,2,2-Tetrachloroethane	10.81	83	23689	2.432	ug/l	97
71) 1,2,3-Trichloropropane	10.85	75	9103	1.310	ug/l #	55
72) Bromobenzene	10.81	156	25469	2.125	ug/l	99
73) n-propylbenzene	10.93	120	39556	2.834	ug/l	91
74) 2-Chlorotoluene	11.01	126	16902	1.410	ug/l	92
75) 1,3,5-Trimethylbenzene	11.11	105	58589	1.373	ug/l	99
76) 4-Chlorotoluene	11.12	126	36093	2.914	ug/l	99
77) tert-Butylbenzene	11.44	119	138370	3.423	ug/l	94
78) 1,2,4-Trimethylbenzene	11.49	105	38586	0.877	ug/l	100
79) sec-Butylbenzene	11.66	105	170189	3.034	ug/l	100
81) p-Isopropyltoluene	11.81	119	39205	0.836	ug/l	100
82) 1,3-Dichlorobenzene	11.77	146	70009	2.920	ug/l	100
84) n-Butylbenzene	12.23	91	117482	2.710	ug/l	99
88) Hexachlorobutadiene	14.04	225	56321	7.002	ug/l	98
89) Naphthalene	14.10	128	136320	5.391	ug/l	99
90) 1,2,3-Trichlorobenzene	14.35	180	64865	4.699	ug/l	99
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU012320\
Data File : VU036539.D
Acq On : 23 Jan 2020 14:34
Operator : JC/MD
Sample : L1134-08DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-UNRVOA-WSDL

Quant Time: Jan 24 02:35:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U012220DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jan 23 08:59:25 2020
Response via : Initial Calibration

