

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012320\
 Data File : VU036540.D
 Acq On : 23 Jan 2020 15:40
 Operator : JC/MD
 Sample : L1134-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Jan 27 09:32:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U012220DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jan 23 08:59:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	6.15	96	36942	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.66	95	12935	0.93	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.00%	
68) 1,2-Dichlorobenzene-d4	12.21	152	13209	0.93	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.00%	
Target Compounds						
					Ovalue	
8) 1,1,2-Trichloro-1,2,2-trif	2.61	101	121085	13.267	ug/l	100
15) Methylene Chloride	3.08	84	2981	0.233	ug/l	94
24) tert-Butyl Alcohol	3.29	59	9768	14.565	ug/l #	72
25) Methyl tert-Butyl Ether	3.41	73	1112753	43.712	ug/l	98
31) Isopropyl Ether	4.05	45	350059	13.206	ug/l	99
32) Ethyl-t-butyl ether	4.56	59	552073	19.905	ug/l	100
33) Tert-Amyl methyl ether	5.99	73	411799	16.215	ug/l	100
73) n-propylbenzene	10.93	120	164495	11.936	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU012320\
Data File : VU036540.D
Acq On : 23 Jan 2020 15:40
Operator : JC/MD
Sample : L1134-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Jan 27 09:32:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U012220DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jan 23 08:59:25 2020
Response via : Initial Calibration

