

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU012320\
 Data File : VU036541.D
 Acq On : 23 Jan 2020 16:07
 Operator : JC/MD
 Sample : L1134-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-EDBCP-WS

Quant Time: Jan 24 02:46:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U012220DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jan 23 08:59:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	6.15	96	38937	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.66	95	13946	0.95	ug/l	0.00
Spiked Amount 1.000			Recovery	=	95.00%	
68) 1,2-Dichlorobenzene-d4	12.22	152	13594	0.90	ug/l	0.00
Spiked Amount 1.000			Recovery	=	90.00%	
Target Compounds						Ovalue
15) Methylene Chloride	3.08	84	3046	0.226	ug/l	98
56) 1,2-Dibromoethane	8.96	107	9440	1.220	ug/l	97
71) 1,2,3-Trichloropropane	10.85	75	11498	1.590	ug/l	75
86) 1,2-Dibromo-3-Chloropropan	13.02	75	2400	1.399	ug/l	84
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU012320\
Data File : VU036541.D
Acq On : 23 Jan 2020 16:07
Operator : JC/MD
Sample : L1134-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-EDBCP-WS

Quant Time: Jan 24 02:46:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U012220DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Jan 23 08:59:25 2020
Response via : Initial Calibration

