

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012325\
 Data File : VU062968.D
 Acq On : 23 Jan 2025 14:25
 Operator : MD/SY
 Sample : VSTD00247
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD002047

Quant Time: Jan 24 00:11:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM012325.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 24 00:10:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	9399	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	10976	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	4387	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	9697	1.545	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.685	65	15795	2.212	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.663	67	10936	1.723	ug/L	-0.02
8) Toluene-d8	7.887	98	33960	2.078	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	10342	1.905	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	11732	1.447	ug/L #	71
6) Trichloroethene	6.520	95	14372	2.104	ug/L	83
9) 1,2-Dibromoethane	8.908	107	7434	1.809	ug/L #	94
12) 1,2,3-Trichloropropane	10.797	75	7781	1.656	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.973	75	1809	1.986	ug/L #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012325\
Data File : VU062968.D
Acq On : 23 Jan 2025 14:25
Operator : MD/SY
Sample : VSTD00247
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD002047

Quant Time: Jan 24 00:11:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM012325.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 24 00:10:41 2025
Response via : Initial Calibration

