

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012325\
 Data File : VU062978.D
 Acq On : 23 Jan 2025 20:05
 Operator : MD/SY
 Sample : Q1137-09
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHAA5

Quant Time: Jan 24 02:22:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM012325.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 24 00:16:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	6905	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	11812	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	3514	0.500	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	2021	0.549	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	110.000%
4) 1,2-Dichloroethane-d4	5.685	65	2908	0.482	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
7) 1,2-Dichloropropane-d6	6.687	67	1931	0.303	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	60.000%
8) Toluene-d8	7.887	98	7371	0.419	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	84.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	2510	0.479	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	96.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012325\
Data File : VU062978.D
Acq On : 23 Jan 2025 20:05
Operator : MD/SY
Sample : Q1137-09
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHAA5

Quant Time: Jan 24 02:22:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM012325.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 24 00:16:16 2025
Response via : Initial Calibration

