

Data File : VU036554.D
Acq On : 24 Jan 2020 14:52
Operator : JC/MD
Sample : VU0124WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK601

Quant Time: Jan 24 23:35:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR012420WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 23:19:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Difluorobenzene	6.29	114	149003	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	137535	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.83	152	64088	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	39664	4.84	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.80%
7) Chloroethane-d5	1.93	69	37206	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.20%
11) 1,1-Dichloroethene-d2	2.59	65	16778	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.40%
20) 2-Butanone-d5	4.70	46	95170	52.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.44%
24) Chloroform-d	5.11	84	74421	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.75	65	39697	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
32) Benzene-d6	5.77	84	166257	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
36) 1,2-Dichloropropane-d6	6.73	67	51925	5.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.40%
41) Toluene-d8	7.93	98	160898	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
43) trans-1,3-Dichloropropene-	8.21	79	19998	4.96	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.20%
46) 2-Hexanone-d5	8.67	63	101072	53.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.96%
56) 1,1,2,2-Tetrachloroethane-	10.78	84	41112	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
66) 1,2-Dichlorobenzene-d4	12.21	152	56944	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU036554.D

Acq On : 24 Jan 2020 14:52

Operator : JC/MD

Sample : VU0124WBL01

Misc : 25.0mL/MSVOA_U/WATER

ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK601

Quant Time: Jan 24 23:35:09 2020

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR012420WMA.M

Quant Title : TRACE VOA SFAM1.0

QLast Update : Fri Jan 24 23:19:07 2020

Response via : Initial Calibration

