

Data File : VU036556.D
Acq On : 24 Jan 2020 15:41
Operator : JC/MD
Sample : VU0124WBL03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK603

Quant Time: Jan 24 23:38:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR012420WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 23:19:07 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Difluorobenzene	6.29	114	144654	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	134877	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.83	152	62094	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	38818	4.88	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.60%
7) Chloroethane-d5	1.93	69	36510	5.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.20%
11) 1,1-Dichloroethene-d2	2.59	65	16672	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.60%
20) 2-Butanone-d5	4.69	46	85579	48.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.66%
24) Chloroform-d	5.11	84	73551	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.75	65	39780	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
32) Benzene-d6	5.77	84	165809	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.20%
36) 1,2-Dichloropropane-d6	6.73	67	51008	5.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.60%
41) Toluene-d8	7.93	98	160456	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
43) trans-1,3-Dichloropropene-	8.21	79	19964	5.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.00%
46) 2-Hexanone-d5	8.67	63	101094	55.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.12%
56) 1,1,2,2-Tetrachloroethane-	10.78	84	41086	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.20%
66) 1,2-Dichlorobenzene-d4	12.21	152	58691	5.44	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU036556.D
Acq On : 24 Jan 2020 15:41
Operator : JC/MD
Sample : VU0124WBL03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK603

Quant Time: Jan 24 23:38:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR012420WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 23:19:07 2020
Response via : Initial Calibration

