

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012423\
 Data File : VU052788.D
 Acq On : 24 Jan 2023 13:47
 Operator : JC/MD
 Sample : 01245-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WSDL

Quant Time: Jan 25 01:36:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Jan 24 05:16:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.112	96	32979	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.629	95	8391	0.765	ug/l	0.00
Spiked Amount 1.000			Recovery	=	76.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	9194	0.815	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.000%	
Target Compounds						
						Qvalue
15) Methylene Chloride	3.044	84	6655	0.716	ug/l	94
27) Chloroform	5.083	83	172640	11.140	ug/l	100
44) Bromodichloromethane	7.102	83	96079	7.878	ug/l	100
55) Dibromochloromethane	8.806	129	79456	9.502	ug/l	100
67) Bromoform	10.285	173	47881	9.703	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012423\
Data File : VU052788.D
Acq On : 24 Jan 2023 13:47
Operator : JC/MD
Sample : 01245-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WSDL

Quant Time: Jan 25 01:36:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012323DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Jan 24 05:16:59 2023
Response via : Initial Calibration

