

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012425\
 Data File : VU062989.D
 Acq On : 24 Jan 2025 13:35
 Operator : MD/SY
 Sample : VIBLK002
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK002

Quant Time: Jan 24 19:04:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 24 19:02:27 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.235	114	101927	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	97346	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.801	152	44177	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	30842	5.463	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.200%
7) Chloroethane-d5	1.904	69	18359	4.043	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.800%
11) 1,1-Dichloroethene-d2	2.557	65	13551	4.634	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.600%
20) 2-Butanone-d5	4.615	46	49556	42.952	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.900%
24) Chloroform-d	5.049	84	58507	4.212	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
26) 1,2-Dichloroethane-d4	5.689	65	30513	4.050	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.000%
32) Benzene-d6	5.714	84	97639	4.103	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
36) 1,2-Dichloropropane-d6	6.676	67	28277	4.286	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.800%
41) Toluene-d8	7.888	98	86043	3.711	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.200%
43) trans-1,3-Dichloroprop...	8.174	79	12926	3.930	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.600%
46) 2-Hexanone-d5	8.627	63	35624	34.441	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	68.880%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	23383	4.232	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.600%
66) 1,2-Dichlorobenzene-d4	12.184	152	30338	4.128	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012425\
Data File : VU062989.D
Acq On : 24 Jan 2025 13:35
Operator : MD/SY
Sample : VIBLK002
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK002

Quant Time: Jan 24 19:04:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 19:02:27 2025
Response via : Initial Calibration

