

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012425\
 Data File : VU062991.D
 Acq On : 24 Jan 2025 14:24
 Operator : MD/SY
 Sample : Q1174-03
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2933

Quant Time: Jan 24 19:05:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 24 19:02:27 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	96810	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.406	117	96009	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	33761	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	30600	5.707	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.200%
7) Chloroethane-d5	1.904	69	18370	4.259	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.200%
11) 1,1-Dichloroethene-d2	2.557	65	13828	4.978	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.600%
20) 2-Butanone-d5	4.608	46	57919	52.854	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.700%
24) Chloroform-d	5.049	84	56869	4.311	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	5.689	65	32685	4.568	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
32) Benzene-d6	5.714	84	99130	4.224	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.400%
36) 1,2-Dichloropropane-d6	6.679	67	30618	4.705	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.200%
41) Toluene-d8	7.888	98	90100	3.941	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.800%
43) trans-1,3-Dichloroprop...	8.171	79	13250	4.085	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.621	63	30587	29.983	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	59.960%
56) 1,1,2,2-Tetrachloroeth...	10.743	84	15912	2.920	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	58.400%#
66) 1,2-Dichlorobenzene-d4	12.184	152	24281	4.323	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012425\
Data File : VU062991.D
Acq On : 24 Jan 2025 14:24
Operator : MD/SY
Sample : Q1174-03
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2933

Quant Time: Jan 24 19:05:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 19:02:27 2025
Response via : Initial Calibration

