

Data File : VU029285.D
Acq On : 25 Jan 2019 13:14
Operator : JC/SP
Sample : K1222-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEJG4

Quant Time: Jan 28 02:30:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jan 28 02:17:15 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	65282	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	65207	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	29888	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	28778	4.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.20%
7) Chloroethane-d5	1.68	69	22640	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.27	63	39663	3.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.40%
20) 2-Butanone-d5	4.21	46	66629	52.27	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.54%
24) Chloroform-d	4.65	84	48306	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	5.31	65	27819	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
32) Benzene-d6	5.34	84	93905	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.33	67	28381	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.56	98	81116	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
43) trans-1,3-Dichloropropene-	7.85	79	12310	4.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.60%
46) 2-Hexanone-d5	8.31	63	59266	52.05	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.10%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	24083	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	32335	5.47	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.40%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU029285.D
Acq On : 25 Jan 2019 13:14
Operator : JC/SP
Sample : K1222-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEJG4

Quant Time: Jan 28 02:30:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jan 28 02:17:15 2019
Response via : Initial Calibration

