

Data File : VU029284.D
Acq On : 25 Jan 2019 12:44
Operator : JC/SP
Sample : VU0125WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK47

Quant Time: Jan 28 02:29:06 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jan 28 02:17:15 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	68610	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	67903	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	31853	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	30701	4.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.60%
7) Chloroethane-d5	1.68	69	22910	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.60%
11) 1,1-Dichloroethene-d2	2.27	63	40457	3.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.20%
20) 2-Butanone-d5	4.22	46	71792	53.59	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.18%
24) Chloroform-d	4.64	84	49643	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
26) 1,2-Dichloroethane-d4	5.31	65	28698	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
32) Benzene-d6	5.34	84	95002	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	6.33	67	30260	5.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.60%
41) Toluene-d8	7.56	98	85903	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
43) trans-1,3-Dichloropropene-	7.85	79	12848	4.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.80%
46) 2-Hexanone-d5	8.31	63	64264	54.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.38%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	24839	4.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	33146	5.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU029284.D
Acq On : 25 Jan 2019 12:44
Operator : JC/SP
Sample : VU0125WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK47

Quant Time: Jan 28 02:29:06 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jan 28 02:17:15 2019
Response via : Initial Calibration

