

Data File : VU029289.D
Acq On : 25 Jan 2019 14:49
Operator : JC/SP
Sample : K1246-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0JJ2

Quant Time: Jan 28 02:52:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jan 28 02:17:15 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	62304	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	63367	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	29739	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	27196	4.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.20%
7) Chloroethane-d5	1.68	69	21824	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.20%
11) 1,1-Dichloroethene-d2	2.27	63	37390	3.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.40%
20) 2-Butanone-d5	4.21	46	67677	55.63	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.26%
24) Chloroform-d	4.65	84	48230	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
26) 1,2-Dichloroethane-d4	5.31	65	28039	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
32) Benzene-d6	5.34	84	89801	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
36) 1,2-Dichloropropane-d6	6.33	67	28198	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.40%
41) Toluene-d8	7.56	98	77752	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
43) trans-1,3-Dichloropropene-	7.85	79	12478	5.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.80%
46) 2-Hexanone-d5	8.31	63	61455	55.53	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.06%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	24839	5.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	32119	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU029289.D
Acq On : 25 Jan 2019 14:49
Operator : JC/SP
Sample : K1246-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0JJ2

Quant Time: Jan 28 02:52:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR012519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jan 28 02:17:15 2019
Response via : Initial Calibration

