

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012623\
 Data File : VU052820.D
 Acq On : 26 Jan 2023 16:52
 Operator : JC/MD
 Sample : 01316-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A44H5

Quant Time: Jan 27 04:34:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM012623WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 27 04:23:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	272988	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	266641	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	147752	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	85511	43.521	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.040%
7) Chloroethane-d5	1.912	69	74938	46.287	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.580%
11) 1,1-Dichloroethene-d2	2.562	63	162456	41.570	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.140%
21) 2-Butanone-d5	4.626	46	118822	97.894	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.890%
24) Chloroform-d	5.057	84	186003	44.357	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.720%
26) 1,2-Dichloroethane-d4	5.694	65	127051	47.451	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.900%
32) Benzene-d6	5.719	84	378148	46.419	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.840%
36) 1,2-Dichloropropane-d6	6.684	67	113397	47.462	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.920%
41) Toluene-d8	7.893	98	347206	46.806	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.620%
43) trans-1,3-Dichloroprop...	8.173	79	59155	49.399	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.800%
47) 2-Hexanone-d5	8.642	63	90675	108.012	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.010%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	172227	48.790	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.580%
66) 1,2-Dichlorobenzene-d4	12.188	152	140017	48.804	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.600%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.604	62	82942	35.562	ug/L	98
8) Chloroethane	1.932	64	73354	50.073	ug/L	96
12) 1,1-Dichloroethene	2.575	96	61428	30.418	ug/L	89
14) Carbon disulfide	2.787	76	290615	47.647	ug/L	100
15) Methyl Acetate	2.948	43	124693	62.428	ug/L	99
16) Methylene chloride	3.041	84	73670	30.667	ug/L	97
18) Methyl tert-butyl Ether	3.359	73	276357	43.065	ug/L	99
19) 1,1-Dichloroethane	3.861	63	79810	20.852	ug/L	96
20) cis-1,2-Dichloroethene	4.661	96	46971	20.159	ug/L	96
27) 1,2-Dichloroethane	5.787	62	201800	59.326	ug/L	98
30) 1,1,1-Trichloroethane	5.308	97	143709	36.641	ug/L	99
34) Trichloroethene	6.536	95	153635	66.596	ug/L	98
37) 1,2-Dichloropropane	6.784	63	52795	23.659	ug/L	99
40) 4-Methyl-2-pentanone	7.796	43	132936	50.565	ug/L	99
42) Toluene	7.964	91	273744	29.428	ug/L	99
44) trans-1,3-Dichloropropene	8.205	75	49525	14.417	ug/L	96
45) 1,1,2-Trichloroethane	8.394	97	90998	39.877	ug/L	98
46) Tetrachloroethene	8.549	164	65423	32.934	ug/L	99
48) 2-Hexanone	8.690	43	109655	49.304	ug/L	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012623\
 Data File : VU052820.D
 Acq On : 26 Jan 2023 16:52
 Operator : JC/MD
 Sample : 01316-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A44H5

Quant Time: Jan 27 04:34:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM012623WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 27 04:23:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


50) 1,2-Dibromoethane	8.918	107	96551	39.312	ug/L	99
51) Chlorobenzene	9.443	112	307679	48.075	ug/L	99
52) Ethylbenzene	9.565	91	607716	61.132	ug/L	99
53) m,p-Xylene	9.687	106	165349	41.086	ug/L	97
54) o-xylene	10.095	106	77829	20.258	ug/L	94
55) Styrene	10.111	104	219325	33.267	ug/L	98
57) 1,1,2,2-Tetrachloroethane	10.780	83	254063	68.796	ug/L	99
59) Bromoform	10.288	173	40674	20.089	ug/L	100
61) Isopropylbenzene	10.481	105	395582	41.454	ug/L	100
64) 1,3-Dichlorobenzene	11.742	146	93265	19.372	ug/L	99
65) 1,4-Dichlorobenzene	11.832	146	196863	39.498	ug/L	98
67) 1,2-Dichlorobenzene	12.208	146	299382	61.479	ug/L	99
68) 1,2-Dibromo-3-chloropr...	12.992	75	60551	73.706	ug/L	99
72) 1,2,3-Trichlorobenzene	14.327	180	110524	34.998	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012623\
Data File : VU052820.D
Acq On : 26 Jan 2023 16:52
Operator : JC/MD
Sample : 01316-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A44H5

Quant Time: Jan 27 04:34:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM012623WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 27 04:23:26 2023
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 35.562 ug/L
 RT: 1.604 min Scan# 81
 Delta R.T. -0.000 min
 Lab File: VU052820.D
 Acq: 26 Jan 2023 16:52




Instrument :
 MSVOA_U
 ClientSampleId :
 A44H5


Tgt Ion: 62 Resp: 82942
 Ion Ratio Lower Upper
 62 100
 64 33.3 22.4 41.6



#8
 Chloroethane
 Concen: 50.073 ug/L
 RT: 1.932 min Scan# 184
 Delta R.T. -0.000 min
 Lab File: VU052820.D
 Acq: 26 Jan 2023 16:52

Tgt Ion: 64 Resp: 73354
 Ion Ratio Lower Upper
 64 100
 66 32.3 24.1 44.7





#12

1,1-Dichloroethene

Concen: 30.418 ug/L

RT: 2.575 min Scan# 3

Delta R.T. -0.003 min

Lab File: VU052820.D

Acq: 26 Jan 2023 16:52

Instrument :

MSVOA_U

ClientSampleId :

A44H5





Tgt Ion: 96 Resp: 61428

Ion Ratio Lower Upper

96 100

61 167.0 114.3 212.3

63 140.3 80.6 149.6



#14

Carbon disulfide

Concen: 47.647 ug/L

RT: 2.787 min Scan# 450

Delta R.T. -0.004 min

Lab File: VU052820.D




Acq: 26 Jan 2023 16:52


Tgt Ion: 76 Resp: 290615

Ion Ratio Lower Upper

76 100

78 8.7 7.1 10.7









#15
Methyl Acetate
Concen: 62.428 ug/L
RT: 2.948 min Scan# 500
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion: 43 Resp: 124693
Ion Ratio Lower Upper
43 100
74 28.2 22.0 33.0



#16
Methylene chloride
Concen: 30.667 ug/L
RT: 3.041 min Scan# 529
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion: 84 Resp: 73670
Ion Ratio Lower Upper
84 100
86 66.0 43.4 80.6
49 104.9 72.4 134.5









#18
Methyl tert-butyl Ether
Concen: 43.065 ug/L
RT: 3.359 min Scan# 61
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion: 73 Resp: 276357
Ion Ratio Lower Upper
73 100
43 18.3 14.2 21.2
57 20.2 16.3 24.5



#19
1,1-Dichloroethane
Concen: 20.852 ug/L
RT: 3.861 min Scan# 784
Delta R.T. -0.004 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52


Tgt Ion: 63 Resp: 79810
Ion Ratio Lower Upper
63 100
65 34.3 22.4 41.6
83 14.3 9.2 17.2








#20
cis-1,2-Dichloroethene
Concen: 20.159 ug/L
RT: 4.661 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Instrument :
MSVOA_U
ClientSampleId :
A44H5







Tgt Ion: 96 Resp: 46971
Ion Ratio Lower Upper
96 100
61 124.0 84.1 156.1
68 0.0 0.0 0.0



#27
1,2-Dichloroethane
Concen: 59.326 ug/L
RT: 5.787 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion: 62 Resp: 201800
Ion Ratio Lower Upper
62 100
98 10.2 7.6 11.4









#30
1,1,1-Trichloroethane
Concen: 36.641 ug/L
RT: 5.308 min Scan# 11
Delta R.T. -0.004 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion: 97 Resp: 143709
Ion Ratio Lower Upper
97 100
99 64.8 51.4 77.0
61 41.6 33.0 49.6



#34
Trichloroethene
Concen: 66.596 ug/L
RT: 6.536 min Scan# 1616
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion: 95 Resp: 153635
Ion Ratio Lower Upper
95 100
97 64.5 44.5 82.7
132 98.4 66.8 124.0
130 104.8 71.9 133.5









#37
1,2-Dichloropropane
Concen: 23.659 ug/L
RT: 6.784 min Scan# 1
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion: 63 Resp: 52795
Ion Ratio Lower Upper
63 100
112 5.2 3.8 5.8



#40
4-Methyl-2-pentanone
Concen: 50.565 ug/L
RT: 7.796 min Scan# 2008
Delta R.T. 0.006 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion: 43 Resp: 132936
Ion Ratio Lower Upper
43 100
58 44.4 35.4 53.2
100 20.7 16.0 24.0





#42

Toluene

Concen: 29.428 ug/L

RT: 7.964 min Scan# 2060

Delta R.T. -0.000 min

Lab File: VU052820.D

Acq: 26 Jan 2023 16:52

Instrument :

MSVOA_U

ClientSampleId :





A44H5

Tgt Ion: 91 Resp: 273744

Ion Ratio Lower Upper

91 100

92 60.0 42.6 79.0



#44

trans-1,3-Dichloropropene

Concen: 14.417 ug/L

RT: 8.205 min Scan# 2135

Delta R.T. -0.000 min

Lab File: VU052820.D




Acq: 26 Jan 2023 16:52


Tgt Ion: 75 Resp: 49525

Ion Ratio Lower Upper

75 100

77 33.9 22.3 41.3









#45
1,1,2-Trichloroethane
Concen: 39.877 ug/L
RT: 8.394 min Scan# 2194
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Instrument :
MSVOA_U
ClientSampleId :
A44H5

Tgt Ion: 97 Resp: 90998




Ion	Ratio	Lower	Upper
97	100		
99	62.5	44.7	83.1
83	83.9	60.3	112.1
85	54.3	38.3	71.1




#46
Tetrachloroethene
Concen: 32.934 ug/L
RT: 8.549 min Scan# 2242
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion:164 Resp: 65423

Ion	Ratio	Lower	Upper
164	100		
129	93.5	64.8	120.3
131	87.0	61.2	113.6
166	129.7	89.2	165.8





#48

2-Hexanone

Concen: 49.304 ug/L

RT: 8.690 min Scan# 21

Delta R.T. 0.006 min

Lab File: VU052820.D

Acq: 26 Jan 2023 16:52

Instrument :

MSVOA_U

ClientSampleId :

A44H5

Tgt Ion: 43 Resp: 109655





Ion Ratio Lower Upper

43 100

58 62.4 48.0 72.0

57 21.6 16.8 25.2

100 16.1 12.7 19.1



#50

1,2-Dibromoethane

Concen: 39.312 ug/L

RT: 8.918 min Scan# 2357

Delta R.T. -0.000 min

Lab File: VU052820.D

Acq: 26 Jan 2023 16:52




Tgt Ion:107 Resp: 96551


Ion Ratio Lower Upper

107 100

109 92.7 65.5 121.6

188 4.6 3.2 4.8









#51
Chlorobenzene
Concen: 48.075 ug/L
RT: 9.443 min Scan# 21
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion:112 Resp: 307679
Ion Ratio Lower Upper
112 100
114 33.2 22.8 42.4
77 56.5 45.9 68.9



#52
Ethylbenzene
Concen: 61.132 ug/L
RT: 9.565 min Scan# 2558
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion: 91 Resp: 607716
Ion Ratio Lower Upper
91 100
106 31.7 22.0 40.8









#53
m,p-Xylene
Concen: 41.086 ug/L
RT: 9.687 min Scan# 21
Delta R.T. -0.004 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion:106 Resp: 165349
Ion Ratio Lower Upper
106 100
91 201.1 137.4 255.2



#54
o-xylene
Concen: 20.258 ug/L
RT: 10.095 min Scan# 2723
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion:106 Resp: 77829
Ion Ratio Lower Upper
106 100
91 213.5 143.4 266.2









#55
Styrene
Concen: 33.267 ug/L
RT: 10.111 min Scan# 21
Delta R.T. 0.003 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion:104 Resp: 219325
Ion Ratio Lower Upper
104 100
78 44.5 31.9 59.2



#57
1,1,2,2-Tetrachloroethane
Concen: 68.796 ug/L
RT: 10.780 min Scan# 2936
Delta R.T. 0.003 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion: 83 Resp: 254063
Ion Ratio Lower Upper
83 100
85 64.9 45.0 83.6
131 10.7 8.6 12.8









#59
Bromoform
Concen: 20.089 ug/L
RT: 10.288 min Scan# 21
Delta R.T. 0.003 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion:173 Resp: 40674
Ion Ratio Lower Upper
173 100
175 47.4 37.7 56.5
254 10.2 8.3 12.5



#61
Isopropylbenzene
Concen: 41.454 ug/L
RT: 10.481 min Scan# 2843
Delta R.T. 0.003 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion:105 Resp: 395582
Ion Ratio Lower Upper
105 100
120 26.9 21.4 32.2
77 15.3 12.2 18.2









#64
1,3-Dichlorobenzene
Concen: 19.372 ug/L
RT: 11.742 min Scan# 3235
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion:146 Resp: 93265
Ion Ratio Lower Upper
146 100
111 39.8 27.8 51.6
148 63.4 45.5 84.5



#65
1,4-Dichlorobenzene
Concen: 39.498 ug/L
RT: 11.832 min Scan# 3263
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52

Tgt Ion:146 Resp: 196863
Ion Ratio Lower Upper
146 100
111 37.4 26.7 49.7
148 62.7 45.6 84.6









#67
1,2-Dichlorobenzene
Concen: 61.479 ug/L
RT: 12.208 min Scan# 311
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52




Instrument :
MSVOA_U
ClientSampleId :
A44H5


Tgt Ion:146 Resp: 299382
Ion Ratio Lower Upper
146 100
111 41.8 33.2 49.8
148 64.1 50.2 75.4



#68
1,2-Dibromo-3-chloropropane
Concen: 73.706 ug/L
RT: 12.992 min Scan# 3624
Delta R.T. -0.000 min
Lab File: VU052820.D
Acq: 26 Jan 2023 16:52


Tgt Ion: 75 Resp: 60551
Ion Ratio Lower Upper
75 100
155 87.6 69.2 103.8
157 114.4 91.2 136.8





#72
 1,2,3-Trichlorobenzene
 Concen: 34.998 ug/L
 RT: 14.327 min Scan# 4039
 Delta R.T. 0.003 min
 Lab File: VU052820.D
 Acq: 26 Jan 2023 16:52

Instrument :
 MSVOA_U
 ClientSampleId :
 A44H5



Tgt Ion:180 Resp: 110524
 Ion Ratio Lower Upper
 180 100
 182 96.2 77.5 116.3
 145 31.3 24.2 36.4

