

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012623\
 Data File : VU052824.D
 Acq On : 26 Jan 2023 18:25
 Operator : JC/MD
 Sample : 01316-12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A44J0

Quant Time: Jan 27 04:35:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM012623WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 27 04:23:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.244	114	271160	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	262132	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	121619	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	86948	44.550	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.100%
7) Chloroethane-d5	1.909	69	75591	47.005	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.020%
11) 1,1-Dichloroethene-d2	2.562	63	134742	34.711	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.420%
21) 2-Butanone-d5	4.623	46	115749	96.005	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.000%
24) Chloroform-d	5.057	84	189027	45.382	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.760%
26) 1,2-Dichloroethane-d4	5.694	65	124872	46.952	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.900%
32) Benzene-d6	5.723	84	379569	47.395	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.800%
36) 1,2-Dichloropropane-d6	6.684	67	111888	47.636	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.280%
41) Toluene-d8	7.893	98	325243	44.600	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.200%
43) trans-1,3-Dichloroprop...	8.173	79	53711	45.624	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.240%
47) 2-Hexanone-d5	8.632	63	82049	99.418	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.420%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	158275	45.609	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.220%
66) 1,2-Dichlorobenzene-d4	12.189	152	121368	51.394	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.780%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012623\
Data File : VU052824.D
Acq On : 26 Jan 2023 18:25
Operator : JC/MD
Sample : 01316-12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A44J0

Quant Time: Jan 27 04:35:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM012623WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 27 04:23:26 2023
Response via : Initial Calibration

