

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012624\
 Data File : VU057686.D
 Acq On : 26 Jan 2024 12:37
 Operator : MD/SY
 Sample : P1229-31
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BB6

Quant Time: Jan 29 00:28:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 29 00:26:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	64744	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	56811	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	24436	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	12920	3.622	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.400%
7) Chloroethane-d5	1.911	69	11932	3.992	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.800%
11) 1,1-Dichloroethene-d2	2.560	65	6373	3.799	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.000%
20) 2-Butanone-d5	4.631	46	35072	52.117	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.240%
24) Chloroform-d	5.055	84	31955	4.413	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	5.698	65	15867	4.822	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.721	84	60272	4.523	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.400%
36) 1,2-Dichloropropane-d6	6.685	67	19119	4.757	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.894	98	54279	4.291	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
43) trans-1,3-Dichloroprop...	8.177	79	6371	3.977	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.631	63	26552	51.950	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.900%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	13138	4.876	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	17391	5.071	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%
Target Compounds						Qvalue
47) Tetrachloroethene	8.550	164	10517	2.664	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012624\
Data File : VU057686.D
Acq On : 26 Jan 2024 12:37
Operator : MD/SY
Sample : P1229-31
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BB6

Quant Time: Jan 29 00:28:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jan 29 00:26:20 2024
Response via : Initial Calibration

