

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012624\
 Data File : VU057680.D
 Acq On : 26 Jan 2024 10:08
 Operator : MD/SY
 Sample : VU0126WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 VBLK137

Quant Time: Jan 29 00:27:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 29 00:26:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	63742	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	57159	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	24497	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	12461	3.548	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	71.000%	
7) Chloroethane-d5	1.910	69	11507	3.910	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	78.200%	
11) 1,1-Dichloroethene-d2	2.563	65	6318	3.826	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.600%	
20) 2-Butanone-d5	4.631	46	33275	50.224	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.440%	
24) Chloroform-d	5.055	84	30731	4.311	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.200%	
26) 1,2-Dichloroethane-d4	5.695	65	14974	4.622	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.400%	
32) Benzene-d6	5.721	84	59351	4.427	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.600%	
36) 1,2-Dichloropropane-d6	6.685	67	18782	4.644	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.800%	
41) Toluene-d8	7.894	98	52235	4.104	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.000%	
43) trans-1,3-Dichloroprop...	8.177	79	7124	4.420	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.400%	
46) 2-Hexanone-d5	8.630	63	26713	51.946	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	103.900%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	13116	4.839	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.800%	
66) 1,2-Dichlorobenzene-d4	12.190	152	16936	4.926	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012624\
 Data File : VU057680.D
 Acq On : 26 Jan 2024 10:08
 Operator : MD/SY
 Sample : VU0126WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK137

Quant Time: Jan 29 00:27:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 29 00:26:20 2024
 Response via : Initial Calibration

