

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012722\
 Data File : VU046869.D
 Acq On : 27 Jan 2022 14:29
 Operator : SY/MD
 Sample : VIBLK023
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK023

Quant Time: Jan 28 01:00:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 28 00:56:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	104567	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	112265	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	50012	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	35158	3.968	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.400%
7) Chloroethane-d5	1.919	69	31320	4.486	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.800%
11) 1,1-Dichloroethene-d2	2.575	65	15117	4.289	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.800%
20) 2-Butanone-d5	4.655	46	125734	47.900	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.800%
24) Chloroform-d	5.067	84	72307	4.648	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.707	65	42070	4.967	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%
32) Benzene-d6	5.732	84	144772	4.388	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
36) 1,2-Dichloropropane-d6	6.694	67	45804	4.624	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.899	98	113893	3.771	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	8.182	79	16693	3.837	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.800%
46) 2-Hexanone-d5	8.639	63	110859	43.034	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.060%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	44517	4.476	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	43158	4.602	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012722\
Data File : VU046869.D
Acq On : 27 Jan 2022 14:29
Operator : SY/MD
Sample : VIBLK023
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK023

Quant Time: Jan 28 01:00:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 28 00:56:24 2022
Response via : Initial Calibration

