

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012722\
 Data File : VU046874.D
 Acq On : 27 Jan 2022 17:00
 Operator : SY/MD
 Sample : N1273-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COHK0

Quant Time: Jan 28 01:01:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 28 00:56:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	105015	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	110020	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	49347	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	32122	3.610	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.200%
7) Chloroethane-d5	1.919	69	29194	4.164	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.200%
11) 1,1-Dichloroethene-d2	2.575	65	14018	3.960	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%
20) 2-Butanone-d5	4.642	46	111988	42.481	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.960%
24) Chloroform-d	5.070	84	68913	4.411	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	5.706	65	38494	4.525	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.400%
32) Benzene-d6	5.732	84	134218	4.151	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.693	67	43073	4.437	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.800%
41) Toluene-d8	7.899	98	107317	3.626	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.600%
43) trans-1,3-Dichloroprop...	8.182	79	15676	3.677	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.600%
46) 2-Hexanone-d5	8.636	63	94565	37.458	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.920%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	39680	4.071	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	41280	4.461	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012722\
Data File : VU046874.D
Acq On : 27 Jan 2022 17:00
Operator : SY/MD
Sample : N1273-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HK0

Quant Time: Jan 28 01:01:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 28 00:56:24 2022
Response via : Initial Calibration

