

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012725\
 Data File : VU063017.D
 Acq On : 27 Jan 2025 15:41
 Operator : MD/SY
 Sample : Q1191-02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A44W7

Quant Time: Jan 29 07:28:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 29 07:27:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	91214	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.406	117	86220	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	43585	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	12892	3.932	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.600%
7) Chloroethane-d5	1.901	69	12438	4.996	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.000%
11) 1,1-Dichloroethene-d2	2.557	65	9557	4.874	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.400%
20) 2-Butanone-d5	4.605	46	40619	42.682	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.360%
24) Chloroform-d	5.049	84	59290	4.552	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
26) 1,2-Dichloroethane-d4	5.689	65	35705	4.802	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	5.714	84	87724	4.452	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
36) 1,2-Dichloropropane-d6	6.676	67	21359	4.258	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.200%
41) Toluene-d8	7.888	98	88526	4.544	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	8.171	79	13383	4.309	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.200%
46) 2-Hexanone-d5	8.621	63	33016	42.606	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.220%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	21398	4.492	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.800%
66) 1,2-Dichlorobenzene-d4	12.184	152	39907	5.115	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012725\
Data File : VU063017.D
Acq On : 27 Jan 2025 15:41
Operator : MD/SY
Sample : Q1191-02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A44W7

Quant Time: Jan 29 07:28:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 29 07:27:08 2025
Response via : Initial Calibration

