

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012725\
 Data File : VU063043.D
 Acq On : 28 Jan 2025 03:14
 Operator : MD/SY
 Sample : Q1195-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E2958

Quant Time: Jan 29 08:19:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 29 08:16:38 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	160448	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.402	117	112639	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	67412	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	17664	3.063	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.200%
7) Chloroethane-d5	1.904	69	15280	3.489	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.800%
11) 1,1-Dichloroethene-d2	2.557	65	10556	3.060	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.200%
20) 2-Butanone-d5	4.605	46	65564	39.166	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.340%
24) Chloroform-d	5.049	84	75892	3.312	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	66.200%#
26) 1,2-Dichloroethane-d4	5.692	65	58543	4.476	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	5.714	84	183051	7.112	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	142.200%#
36) 1,2-Dichloropropane-d6	6.679	67	47652	7.271	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	145.400%#
41) Toluene-d8	7.888	98	106586	4.188	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.800%
43) trans-1,3-Dichloroprop...	8.171	79	23514	5.795	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	116.000%
46) 2-Hexanone-d5	8.621	63	86737	85.678	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	171.360%#
56) 1,1,2,2-Tetrachloroeth...	10.746	84	55966	8.993	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	179.800%#
66) 1,2-Dichlorobenzene-d4	12.183	152	63267	5.243	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012725\
Data File : VU063043.D
Acq On : 28 Jan 2025 03:14
Operator : MD/SY
Sample : Q1195-06
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E2958

Quant Time: Jan 29 08:19:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 29 08:16:38 2025
Response via : Initial Calibration

