

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012725\
 Data File : VU063021.D
 Acq On : 27 Jan 2025 17:21
 Operator : MD/SY
 Sample : Q1192-09
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A44W9

Quant Time: Jan 29 07:29:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 29 07:27:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	111296	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	101656	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	51081	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	17685	4.420	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.901	69	12672	4.172	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.400%
11) 1,1-Dichloroethene-d2	2.554	65	11487	4.801	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.000%
20) 2-Butanone-d5	4.608	46	55412	47.720	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.440%
24) Chloroform-d	5.049	84	68443	4.307	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	5.689	65	41008	4.520	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.400%
32) Benzene-d6	5.714	84	105617	4.547	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
36) 1,2-Dichloropropane-d6	6.679	67	29536	4.993	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.800%
41) Toluene-d8	7.888	98	100289	4.366	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	8.171	79	15820	4.320	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.400%
46) 2-Hexanone-d5	8.621	63	44156	48.329	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.660%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	27076	4.821	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.400%
66) 1,2-Dichlorobenzene-d4	12.184	152	41249	4.511	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012725\
Data File : VU063021.D
Acq On : 27 Jan 2025 17:21
Operator : MD/SY
Sample : Q1192-09
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A44W9

Quant Time: Jan 29 07:29:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 29 07:27:08 2025
Response via : Initial Calibration

