

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012725\
 Data File : VU063027.D
 Acq On : 27 Jan 2025 19:49
 Operator : MD/SY
 Sample : Q1192-08
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A44W3

Quant Time: Jan 29 07:30:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 29 07:27:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	106567	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	87302	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	50613	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18555	4.844	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.800%
7) Chloroethane-d5	1.904	69	14768	5.077	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.600%
11) 1,1-Dichloroethene-d2	2.557	65	12314	5.375	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	107.600%
20) 2-Butanone-d5	4.605	46	67541	60.747	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.500%
24) Chloroform-d	5.049	84	68655	4.512	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.688	65	42010	4.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.800%
32) Benzene-d6	5.714	84	106245	5.326	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.600%
36) 1,2-Dichloropropane-d6	6.679	67	31321	6.166	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	123.400%
41) Toluene-d8	7.888	98	100122	5.076	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
43) trans-1,3-Dichloroprop...	8.171	79	16546	5.261	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.200%
46) 2-Hexanone-d5	8.621	63	51670	65.852	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	131.700%#
56) 1,1,2,2-Tetrachloroeth...	10.746	84	30303	6.283	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	125.600%#
66) 1,2-Dichlorobenzene-d4	12.187	152	43188	4.767	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012725\
Data File : VU063027.D
Acq On : 27 Jan 2025 19:49
Operator : MD/SY
Sample : Q1192-08
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A44W3

Quant Time: Jan 29 07:30:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR012725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 29 07:27:08 2025
Response via : Initial Calibration

