

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012821\
 Data File : VU042213.D
 Acq On : 28 Jan 2021 20:29
 Operator : SY/MD
 Sample : M1126-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Jan 29 10:31:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jan 29 09:59:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.124	96	13331	1.00	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	5047	0.98	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.00%	
68) 1,2-Dichlorobenzene-d4	12.204	152	5059	0.96	ug/l	0.00
Spiked Amount 1.000			Recovery	=	96.00%	
Target Compounds						
8) 1,1,2-Trichloro-1,2,2-...	2.591	101	70669	20.90	ug/l	Qvalue 100
24) tert-Butyl Alcohol	3.218	59	6280	33.80	ug/l #	72
25) Methyl tert-Butyl Ether	3.378	73	143886	15.94	ug/l	99
31) Isopropyl Ether	4.012	45	482261	39.72	ug/l	100
32) Ethyl-t-butyl ether	4.523	59	390548	35.78	ug/l	100
33) Tert-Amyl methyl ether	5.960	73	184943	18.11	ug/l	100
73) n-propylbenzene	10.915	120	164902	34.63	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012821\
Data File : VU042213.D
Acq On : 28 Jan 2021 20:29
Operator : SY/MD
Sample : M1126-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Jan 29 10:31:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jan 29 09:59:59 2021
Response via : Initial Calibration

