


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012821\
 Data File : VU042201.D
 Acq On : 28 Jan 2021 11:41
 Operator : SY/MD
 Sample : BFB
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U012821DW.M
 Title : METHOD 524.2 VOLATILES DRINKING WATER
 Last Update : Fri Jan 29 09:59:59 2021



AutoFind: Averaged scan 2880 to 2912; Bkg corrected with scan 2879

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	2573	PASS
75	95	30	60	53.7	6540	PASS
95	95	100	100	100.0	12179	PASS
96	95	5	9	6.6	807	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	83.0	10107	PASS
175	174	5	9	7.2	728	PASS
176	174	95	101	97.3	9839	PASS
177	176	5	9	6.6	650	PASS